

CALIBRATION CERTIFICATE

This is to certify that the following Biochemistry Analyser has been tested on the below aspects.

Customer Name : BHARATH X RAY & LAB
Location : 715/13, Thrichy Main Road
 NAMAkkAL, TAMILNADU
Instrument : TURBOCHEM100 (4621)CHEMISTRY AUTO ANALYZER
Installation Date : 12-05-2017
SerialNo : 4621-1223
Calibrated On : 01-06-2022
Next Calibration Due On : 31-05-2023

FILTERS :

The Photometric Voltages should be within 2V to 10V

Filters (nm)	Voltages (v)
340	3.06
405	3.92
505	4.49
545	4.44
580	4.63
630	4.99

Pipetting Check

Requirement : The C.V. should be less than 2% for Medium & Large, less than 3% for Small.

Self Test - Passed

Type	Target	Mean	SD	CV
SMALL	0.35	0.41	0.01	2.47
MEDIUM	1.73	1.75	0.01	0.66
LARGE	2.67	2.56	0.02	0.90

Incubation Temperature

Range : 37°C + 0.5

On Board display the Temperature is 37.0 degree stable

On complete testing of the unit we found that the unit is working satisfactorily and used to carry out the necessary functions.

THIS IS AN ELECTRONICALLY GENERATED STATEMENT HENCE SIGNATURE NOT REQUIRED