



ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SECTOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com CC-3427



Calibration Certificate

CERTIFICATE NO.	: 221122B/N-106	Page	: 1 of 1	
CALIBRATED ON.	: 22 / 11 / 2022	Ref. Std.	: ISO:8655:6	
CALIBRATION DUE.	: 21 / 05 / 2023	C.P. No.	: ACS/CP/08	
SRF No. & DATE	: 221122B/N Dt. 22.11.2022	Phy. Inspection	: O.K.	
ULR No. & ISSUE DATE	: CC342722000001009F Dt. 26.11.2022	Cal. At	: Site / Lab	
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. H-2/150, Narmada Nagar, Mangla Chowk Bus Stop, Near Heaven Building, Bilaspur, Chattishgarh			
ITEM CALIBRATED	: Micro Pipette			
	Make : Transasia	Range	: (10 to 100) μ l	
	L. Count :	ID No.	: RCL/SER/PIP-06	
	Location : Serology	S. No.	: YE212AS0143211	
ENVIRONMENT CONDITION:	Temp. { 25°C \pm 3°C} R.H. {40% to 60%}			
STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Weighing Balance	ACS/DB/02/ D449927064	BlueBox	BBT/223/JAN/22	06/01/2023
RESULTS : Mechanical Discipline				
Nominal Capacity (μ l)	Graduation (μ l)	Nominal Value (μ l)	Volume at 20 °C (μ l)	Error (μ l)
		10	10.15	(-) 0.15
100		50	50.24	(-) 0.24
		100	100.40	(-) 0.40

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is $\pm 0.02 \mu$ l at coverage factor $k=1.96$.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

