

Erba Lachema s.r.o.



Brno, 10.11.2016

To,
Redcliffe Lifetech Pvt Ltd.
Bilaspur, Chhattisgarh

Instrument Name: LAURA
Instrument Serial no: 0964-1-202201

We Erba Lachema SRO with registered office in Czech Republic, Karasek 2219/1D, 621 00 BMO, represented by its Production manager Veronika Pleskova here by confirm as Producer of LAURA instrument that this device is firmly calibrated by the manufacturer and does not require any regular calibration by the user, correct calibration of instrument is controlled by grey strips provided initially together with the instrument.

P.S.
A handwritten signature in blue ink, appearing to read "P.S.", followed by a large, stylized signature.

RNDr. Pavel Suchan Ph.D.

Global Product Manager, Urine Analysis

Erba Lachema s.r.o.

Karásek 1d, Brno 62100, Czech Republic



DEKAPHAN LAURA

Seq.No: 0009

URINORM N

2023.01.14

08:02 AM

.....

BLD	NEG
BIL	NEG
UBG	NORM
KET	NEG
GLU	NEG
PRO	NEG
pH	5
NIT	NEG
LEU	NEG
SG	1.025

QC TEST: OK
<Admin>

DEKAPHAN LAURA

Seq.No: 0002

URINORM P

2023.01.14

07:55 AM

.....

*BLD	250	Ery/u1
*BIL	3	mg/dl
*UBG	1	mg/dl
*KET	16	mg/dl
*GLU	1000	mg/dl
*PRO	100	mg/dl
pH	9	
*NIT	POS	
LEU !!	NEG	
SG	1.000	