



Hemoglobin (Hb1Ac)

ALL METHOD REPORT

Cycle - 11/2022
Round No - 4

Date: 20/10/2022

Lab Code: 2617

Parameters	Units	No.of. Participants	Group Mean	Standard deviation (SD)	Uncertainty of Assign Values	Range (± 2 SD)	Your Value	Standard Deviation Index (SDI)
Hemoglobin (Hb A1C)	%	62	5.6	0.6	0.09	4.5-6.7	5.7	0.2

Interpretation of SDI:

SDI Value(+/-)	0 - 0.5	0.6 - 0.9	1.0 - 2.0	2.1 - 2.9	≥ 3
Interpretation	Excellent Performance	Good Performance	Acceptable Performance	Marginal Performance Needs Improvement	Unacceptable Performance Needs Urgent action

Legends	(*) Excluded From Group	{ } Not Reported	(#)Late Result	(\$)Reported in
---------	-------------------------	------------------	----------------	-----------------

Chief Coordinator

Dr. Sanjay Mehrotra
Dr. Sanjay Mehrotra

Checked By: *[Signature]*

Prepared by: SM *[Signature]*

Programme Director

Dr. Bandana Mehrotra
Dr. Bandana Mehrotra

Page 1 of 1

