


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)
Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

CERTIFICATE NO.	: 221024A/N-114	Page	: 1 of 1
CALIBRATED ON.	: 25 / 10 / 2022	Ref. Std.	: CG-11
CALIBRATION DUE.	: 24 / 10 / 2023	C.P. No.	: ACS/CP/07
SRF No. & DATE	: 221024A/N Dt. 24.10.2022	Phy. Inspection	: O.K.
ULR No. & ISSUE DATE	: CC34272200000943F Dt. 21.11.2022	Cal. At	: Site / Lab
CALIBRATED FOR	: M/s Redcliffe Lifetech Pvt. Ltd. CG-223, Salt Lake, Sector-2, Kolkata 700091		
ITEM CALIBRATED	: Digital Temperature Indicator		
	Make : RoHS	Range	: (-50 to 110) °C
	L. Count :	ID No.	: RCL/BIO/D.Temp-02
	Location : Clinical Biochemistry	S. No.	: N.A
ENVIRONMENT CONDITION:	Temp. { 25°C ± 3°C} R.H. {40% to 60%}		

STANDARD INSTRUMENTS USED :				
INSTT. NAME	ID No. / S. No.	TRACEABILITY	CERTIFICATE No.	DUE DATE
Digital Calibrator	ACS/UC/01 / 300561	Anshaankan	AIPL/2021/0784	11/11/2022

RESULTS : ET Discipline			
* U. C. INSTT. READS (°C)	STANDARD READING (°C)	ERROR (°C)	
2.02	2	0.02	
4.04	4	0.04	
6.08	6	0.08	
8.10	8	0.10	
(average of readings)			

REMARKS : The Expanded uncertainty of measurement at 95% confidence level is ± 1.00°C at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Confirmity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

