


ADVANCE CALIBRATION SERVICES

(ISO : 17025 NABL ACCREDITED LAB.)

'C' - 23, SETOR-10, NOIDA - 201301 (U.P.)

Mob. : +91-9891121066, 7982584097 • E-mail : advance.cal@gmail.com



Calibration Certificate

| | | | | |
|------------------------------------|---|------------------------|------------------------|-------------------------|
| CERTIFICATE NO. | : 221024A/N-107 | Page | : 1 of 1 | |
| CALIBRATED ON. | : 24 / 10 / 2022 | Ref. Std. | : DKD R-5-7 | |
| CALIBRATION DUE. | : 23 / 10 / 2023 | C.P. No. | : ACS/CP/11 | |
| SRF No. & DATE | : 221024A/N Dt. 24.10.2022 | Phy. Inspection | : O.K. | |
| ULR No. & ISSUE DATE | : CC342722000000936F Dt. 21.11.2022 | Cal. At | : Site / Lab | |
| CALIBRATED FOR | : M/s Redcliffe Lifetech Pvt. Ltd. CG-223, Salt Lake, Sector-2, Kolkata 700091 | | | |
| ITEM CALIBRATED | : Digital Thermo Hygrometer | | | |
| | Make | : Digital Series | Range | : (0 to 50)°C, (10-99)% |
| | L. Count | : | ID No. | : RCL/HEM/HGM-01 |
| | Location | : Haematology | S. No. | : N.A |
| ENVIRONMENT CONDITION: | Temp. { 25°C ± 3°C } R.H. { 40% to 60% } | | | |
| STANDARD INSTRUMENTS USED : | | | | |
| INSTT. NAME | ID No. / S. No. | TRACEABILITY | CERTIFICATE No. | DUE DATE |
| RH & Temp. Sensor with Indicator | ACS/RHTIS/01/ 20365872 | DVG | CC210122000001272F | 15/07/2023 |
| RESULTS | : Thermal Discipline | | | |
| * U. C. INSTT. READS | STANDARD READING | ERROR | | |
| (°C) | 15.02 | 14.99 | 0.03 | |
| | 25.04 | 24.99 | 0.05 | |
| | 40.06 | 39.98 | 0.08 | |
| (%) | 30.04 | 29.99 | 0.05 | |
| | 60.06 | 59.98 | 0.08 | |
| | 90.08 | 89.97 | 0.11 | |

REMARKS

: The Expanded uncertainty of measurement at 95% confidence level is ± 0.8°C, 1.9% at coverage Factor k=1.96.

*U.C. : Under Calibration.

- Instrument has been Calibrated against laboratory Standard instruments whose values are traceable to National Standard as mention in above.
- The calibration results reported in this certificate are valid at the time of and under the stated conditions of measurement for a particular sample identified above.
- This Certificate should not be reproduced in full or in an abstract form without obtaining prior written permission from Advance Calibration Services.
- The Calibration Certificate is not to be used for any legal purpose and shall not be produced in the court of law.
- Instruments has been Calibrated only for Scientific, in house, Testing and Industrial use & should not be used for trade / commercial use.
- Conformity statement not provided since not required.

Prepared by

....End of Calibration Certificate....

