

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Pillanallur
 Lab No 7451
 Month November
 Year 2022
 Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)


Date of Result Entered : 22/11/2022

Date of Report Published : 06/12/2022

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	80 mg/dL	-0.41	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	7.7 mg/dL	1.45	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	111 mg/dL	-0.25	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links


 22/11/2022
 Govt. Block PHC, Pillanallur
 District: Tirunelveli
 State: Tamil Nadu