

VIEW LAB MONTHLY SUMMARY

Lab Name Govt. Block PHC, Pillanallur

Details About Robust Analysis

Lab No 7451

Detail About Monthly Summary

Month October

Detail about SDI

Year 2022

Constituent Group Chemistry III

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)


Date of Result Entered : 19/10/2022

Date of Report Published : 03/11/2022

1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1837	346.52	7.17	24.84	365 mg/dL	0.74	1.16
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1831	1.36	10.34	0.14	1.4 mg/dL	0.28	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	109.51	6.27	6.87	107 mg/dL	-0.37	0.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within +1.01 to +2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links


Dr. J. Jeyaraj
 Director
 Government Block PHC, Pillanallur
 Pillanallur, Villupuram District
 Tamil Nadu - 605 004