



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - SEPTEMBER 2022



PC-1024

Lab Name	Govt. Block PHC, Pillanallur	Lab No	7451
Constituent Group	Chemistry III	Date of Result Entered :	20/09/2022
PT item	Lyophilized Serum	Date of Report Published :	05/10/2022

Sl. No.	Parameter	Method	Instrument	Result	Target	Delta	Delta 2	Delta 3	Delta 4	Delta 5
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1846	223.49	6.67	14.90	219 mg/dL	-0.30	0.69
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1834	3.49	9.05	0.32	3.8 mg/dL	0.98	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1798	112.08	5.53	6.20	111 mg/dL	-0.17	0.29

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Pillanallur
 THE BLOCK MEDICAL OFFICER, CLINICAL LABORATORY, NAMAKKAL DISTRICT.
 Pillanallur
 TAMILNADU637403

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****