

**CENTRIFUGE CALIBRATION REPORT**

DEPARTMENT <u>Accession</u>	INSTRUMENT NAME <u>Centrifuge</u>	INSTRUMENT ID <u>ACC/001</u>	MODEL NO <u>R8C</u>	REPORT NO <u>CAL/CEG/02</u>
Cal Done Date <u>18/01/2023</u>	Cal Due Date <u>18/07/2023</u>			
Ownership SRL <input checked="" type="checkbox"/>	LOAN <input type="checkbox"/>		RR <input type="checkbox"/>	
Acceptable rpm/time limit $\pm 5\%$	Operating Range From <u>1000</u> rpm To <u>4000</u> rpm			

Sr. No.	Set Value	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.
1	1000	1006.4	2.51	5.935	6.325
2	2000	2007.0	2.55	6.635	6.338
3	4000	4013.4	2.302	13.235	6.262

Sr. No.	Set Value	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.
1	60	60.32	0.13	0.086	0.164
2	300	300.48	0.164	0.089	0.247
3	600	600.88	0.13	0.293	0.372

Sr. No.	Set Value °c	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing the level of confidence of approx. 95%.

Calibration status:  Passed  Failed

TEST EQUIPMENT NAME	Test equipment ID	Traceable Report No
TACHOMETER	<u>ENG/DTM-01</u>	<u>MTL/SRL2/RO2/01-23</u>
TIMER	<u>ENG/SW-01</u>	<u>MT4/SRL2/RO6/01-23</u>
THERMOMETER		

The above mentioned instrument has been calibrated using standards with accuracy guidelines traceable to the National Institute of Standards & Technology (NIST).

Engineer's Name <u>Mirajul Islam</u>	Report Reviewed by
Signature/Date <u>18/1/23</u>	Signature/Date <u>D. Saha</u>

**SECTION HEAD SIGN:**

Doc No.: <u>ENG QC02(316)</u>	Centrifuge Calibration Report			Page 1 of 1
Issue No.: <u>01</u>	Issue Date: <u>01.12.2021</u>	Amend No.: <u>00</u>	Amend Date: <u>00</u>	<u>Amph</u>
Issued By: <u>Mr. Surendra Kumar Gupta (DGM-QA)</u>				



# MEASURE TECHNO LAB

2, B.T. ROAD (JAYANTI CINEMA COMPLEX), BARRACKPORE,  
KOLKATA - 700120, W.B.

Phone : 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974,  
LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in



NABL ACCREDITED LABORATORY  
Certificate No. CC - 2343

## CALIBRATION CERTIFICATE OF DIGITAL TACHOMETER

CALIBRATION CERTIFICATE NO.: MTL / SRL / R02 / 01 - 23

ULR - CC254523000001407F

Page: 1 of 1

1.0 Service Request No.: MTL / 14 / 01 / 22 - 23

1.1 Issued to: M/s. SRL Limited,  
P.S. Srijan Tech Park, Ground Floor,  
DN - 52, Sector V, Salt Lake,  
Kolkata - 700091, W.B.

1.2 Description & Identification of Item to be Calibrated:	a) Name:	Digital Tachometer	b) Code No.:	ENG / DTM - 01
	c) Sl. No.:	AB.16577	d) Make:	Lutron
	e) Model / Type:	DT - 2236	f) Range:	5 rpm to 99999 rpm
	g) Sensor:	N.A.	h) Resolution:	0.1 rpm upto 999.9 rpm then 1 rpm
	i) End User:	Engineering Department	j) Accuracy:	N.S.
	k) Calibration done at:	On Site / <input checked="" type="checkbox"/> In House		

1.3 Date of receipt of item : 13-01-23

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 14-01-23

1.6 Recommended date of next calibration : 14-01-24

1.7 Date of Issue : 16-01-23

1.8 Environmental Conditions During Calibration: Temperature: 20 °C ± 2 °C  
Humidity: 50 % RH ± 10 % RH

1.9 Method of Calibration: SOP / RPM / 01

### 2.0 Traceability :

a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.

b) The following standards / Equipment have been used.

i) Digital Tachometer Cal. Certificate No. JRPM - CCTR - A&S - 2022 - 0025 (JRPM, Chennai) (Cal. Date: 09/07/22, Due Date: 08/07/23)

### 2.1 Result :

#### Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value on DUC rpm	Std. Value Average of Five Readings rpm	Error rpm	Measurement Expanded Uncertainty ± rpm
1.	Rotation 5 rpm to 99999 rpm	100.1	100.11	-0.01	0.28 upto 100 rpm then 1.28 upto 10000 rpm then 14 upto 30000 rpm then 10.43
2.		499.9	499.85	0.05	
3.		1000	1000.5	-0.5	
4.		4999	4999.6	-0.6	
5.		9987	9986	1	
6.		19999	19995	4	
7.		29988	29992	-4	
8.		59983	59980	3	
9.		89485	89487	-2	

Remarks: i) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.

ii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.

iii) Error = DUC Reading - Standard Reading.

DUC - Device Under Calibration

N.S. - Not Specified

N.A. - Not Applicable

Std. - Standard

rpm - Rotation Per Minute

#### Opinions and Interpretations

Calibrated	✓	Accepted / Valid for use
Limited Use		Rejected / Out of use

Calibration Engineer  
Measure Techno Lab  
R.K. Singh  
Calibration Engineer

Checked / Approved by

