



Rapid Diagnostic Pvt. Ltd.



Calibrator Calibration

Instrument:	Model:	Serial No.:	Date of Cal.			
5 Part Haematology Analyzer	BC -6200	TW-03000947	07.01.2023			
Mode: OV-WB	Calibrator Lot No. PLUSE0123		Exp.Date:05.02.2023			
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	8.92	4.61	13.6	92.4	255	8.92
1.	7.25	4.85	14.3	92.6	238	
2.	7.28	4.87	14.4	92.5	244	
3.	7.44	4.83	14.3	92.4	250	
4.	7.57	4.89	14.4	92.3	240	
5.	7.50	4.93	14.4	92.1	252	
6.	7.27	4.85	14.4	92.2	260	
Mean	1.846	4.87	14.3	92.4	247	
CV(%)	1.846	0.766	0.28	0.20	3.3	
New Cal. Factor (%)	120.79	94.61	94.84	100.05	103.06	
Old Cal. Factors (%)	103.67	92.06	94.87	100.38	93.29	
Old TC1 (%)						101.53
New TC1 (%)						86.19

For Rapid Diagnostic Pvt. Ltd.


R. Sidharth
Service Engineer


N. Mahender Reddy
Zonal Service Manager

B-82, G.T. Karnal Road Industrial Area, Delhi-110 033, Ph.: 91-11-45201100, 27113030, Fax: 91-11-27413437, 45201198
e-mail: drjindal@rdgc.in, website: <http://www.rapiddiagnostic.in>



Rapid Diagnostic Pvt. Ltd.



ISO 9001:2008
CERTIFIED COMPANY

Calibration Certificate

Customer Name	Telangana Diagnostic Hub Nalgonda
Address	Area Hospital, Nalgonda-508001
Contact Person	Sumaiya
Contact No.	630575104
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000947
Instrument Status	Under Warranty
Calibration Done On	07.01.2023
Next Calibration Due On	06.01.2024

For Rapid Diagnostic Pvt. Ltd.

R. Sidharth
Service Engineer

N. Mahender Reddy
Zonal Service Manager

B-82, G.T. Karnal Road Industrial Area, Delhi-110 033, Ph.: 91-11-45201100, 27113030, Fax: 91-11-27413437, 45201198
e-mail: drjindal@rdgc.in, website: <http://www.rapiddiagnostic.in>