



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name **GURU MICRO LABORATORY** Lab No **16065**
 Constituent Group **Chemistry II** Date of Result Entered : **18/11/2022**
 PT item **Lyophilized Serum** Date of Report Published : **06/12/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Agape	348	187.48	8.36	15.67	191.3 mg/dL	0.24	1.68
2	UREA II	Urease UV / GLDH II	Agape	308	51.71	10.04	5.19	46.9 mg/dL	-0.93	0.59
3	CREATININE II	Jaffes Kinetic-Alkaline Picrate II	Agape	249	2.36	8.64	0.20	2 mg/dL	-1.76	0.03
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Agape	256	3.02	11.11	0.34	3.4 mg/dL	1.13	0.04
5	T-PROTEIN II	Biuret - Colorimetric II	Biosystems	260	5.55	7.98	0.44	4.9 g/dL	-1.47	0.05
6	ALBUMIN II	BCG - Colorimetric II	Biosystems	255	3.32	9.76	0.32	2.9 g/dL	-1.30	0.04
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Agape	263	5.22	15.89	0.83	5 mg/dL	-0.27	0.10
8	CHOLESTROL II	CHOD-PAP II	Agape	341	107.20	12.47	13.37	113 mg/dL	0.43	1.45
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Agape	240	174.08	9.80	17.06	172 mg/dL	-0.12	2.20

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

GURU MICRO LABORATORY

OPPOSITE TO DEVI PHARMACY, NEAR POLICE STATION, SALEM MAIN ROAD, KOMARAPALAYAM,

NAMAKKAL

TAMILNADU638183

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss

Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.

Data in CMC EQAS reports is confidential

CMC EQAS does not sub contract any components

***** End of Report *****