DIVINE DIAGNOSTICS

SHOP NO 51, GROUND FLOOR, RAJAT SANKUL NEAR S T STAND GANESH PETH NAGPUR 440018

Date 11-12-2022

Calibration Certificate

This is to certify that Haematology Analyzer BC 3000 plusInstrument Sr. No. RJ34114645 Installed at ASHA HOSPITAL, KAMPTEE has been calibrated by our engineer on 11 Dec -2022 with appropriate and sophisticated tools. Below parameters are been calibrated:

This calibration includes preventive maintenance, calibration of all moving parts.

TEST ITEMS AND RESULT

Appearance	OK
Test Speed	OK
Suction samp le program	OK
Valve and optical application	OK
Motor & valve	OK
Gain ad jus tment and red, white MCV test	OK
Samp le Volu me	OK
Diamet er Orifice	OK
Hydra uli c Unit	OK
Power Consumption	250V 4A
Fuse check	OK
Power Supply and Earthing	230 Volts and Grounding less then SV


This Calibration Certificate is Valid Up to 10 -Dec -2023

Conclusion: All parameter are within range. The Machine is working properly.

Engineer Name: Rakesh Ramteke

Engineer Sign: ^

DIVINE DIAGNOSTICS
Shop No. 54, Ground Floor.
Rajat Sankul, Near S.T. Bus StaGanesh Peth, Nagpur- 440018



DL NO 20B/125072 21B/125073

DIVINE DIAGNOSTICS

SHOP NO 51 , GROUND FLOOR , RAIAT SANKUL NEAR S T STAND GANESH PETH NAGPUR 440018

To

The Director

Asha Hospital

Kamptee

Subject - Service Contract

Dear Sir

This I is to inform you that your Heamatology Analyzer BC 3000+ will be under service contract from 10th Dec- 2022 to 09th Dec 2023 against purchase Meril Brand of reagents from Divine Diagnostics Nagpur

Our Engineer will continue to give free service and application support to the instrument and take care of routine maintenance during contract period

Spare parts required due to any major breakdown will be charged as per applicable rates

Thanks

For Shop No. 51 Ground Floor.
Rajat Sankul, Near S.T. Bus St. ...,
Ganesh Peth, Nagpur- 440018

Divine Diagnostics