

VIEW LAB MONTHLY SUMMARY

Lab Name	Govt. PHC, Perumuchi	Details About Robust Analysis
Lab No	13002	Detail About Monthly Summary
Month	August	Detail about SDI
Year	2022	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 17/08/2022

Date of Report Published : 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1833	104.52	7.14	7.46	92 mg/dL	-1.68	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1780	6.46	8.21	0.53	4.9 mg/dL	-2.94	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1717	115.04	5.56	6.39	114 mg/dL	-0.16	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

Modify Methods	Submit Lab Results	View Lab Monthly Summary
Change Password	Home	Sign-out

