

## VIEW LAB MONTHLY SUMMARY

Lab Name	GOVT PRIMARY HEALTH CENTER	<a href="#">Details About Robust Analysis</a>
Lab No	13002	<a href="#">Detail About Monthly Summary</a>
Month	July	<a href="#">Detail about SDI</a>
Year	2022	
Constituent Group	Chemistry III	

Click on the analyte to view Graphical Data

[All Analyser Result](#)[Print](#)

Date of Result Entered : 16/07/2022

Date of Report Published : 05/08/2022

	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
					CV	SD			
1	GLUCOSE III	GOD-POD III	1864	295.51	7.14	21.09	307 mg/dL	0.54	0.98
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	1839	3.40	9.21	0.31	3.6 mg/dL	0.64	0.01
3	CHOLESTROL III	CHOD-PAP III	1803	117.79	6.19	7.29	108 mg/dL	-1.34	0.34

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## Links

<a href="#">Modify Methods</a>	<a href="#">Submit Lab Results</a>	<a href="#">View Lab Monthly Summary</a>
<a href="#">Change Password</a>	<a href="#">Home</a>	<a href="#">Sign-out</a>

