

VIEW LAB MONTHLY SUMMARY

Lab Name GOVT PRIMARY HEALTH CENTER

Lab No 13002

Month March

Year 2022

Constituent Group Chemistry III

[Details About Robust Analysis](#)

[Detail About Monthly Summary](#)

[Detail about SDI](#)

[Click on the analyte to view Graphical Data](#)

[All Analyser Result](#)

[Print](#)

Date of Result Entered : 19/03/2022

Date of Report Published : 06/04/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants CV	SD	Year Value	SDI	IP
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1766	151.83	7.28	11.06	142 mg/dL	-0.89	0.53
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1708	1.20	11.54	0.14	1.3 mg/dL	0.72	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1655	115.16	6.70	7.71	114 mg/dL	-0.15	0.38
<b>SDI Range</b>				<b>Interpretation</b>						
Within -1.00 to +1.00				Excellent.						
Within ±1.01 to ±2.00				Good.						
Within ±2.01 to ±2.99				Accept with caution. Warning Signal.						
Beyond ±3.0				Unacceptable performance. Action Signal.						

[Links](#)