



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - AUGUST 2022



Lab Name Govt. UPHC, Kallukulam
Constituent Group Chemistry III
PT item Lyophilized Serum

Lab No 9388
Date of Result Entered : 20/08/2022
Date of Report Published : 05/09/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1833	104.52	7.14	7.46	108 mg/dL	0.47	0.35
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1780	6.46	8.21	6.53	6.1 mg/dL	-0.88	0.03
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1717	115.04	5.56	6.39	117 mg/dL	0.31	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UPHC, Kallukulam
KALLUKULAM, N.K. ROAD
Thanjavur
TAMILNADU 613001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2263102

Pamela Christudoss

Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

flust
DR. A.S. MUTHU KUMAR
MEDICAL OFFICER
UPHC, KALLUKULAM,
THANJAVUR.
9865656519

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential