



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - JUNE 2022



Lab Name Govt. UPHC, Kallukulam

Lab No 9388

Constituent Group Chemistry III

Date of Result Entered : 17/06/2022

PT item Lyophilized Serum

Date of Report Published : 07/07/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	S
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1828	87.43	7.74	6.77	85 mg/dL	-0.36	0.32
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1769	1.99	8.26	0.16	1.8 mg/dL	-1.16	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1737	118.45	6.70	7.93	120 mg/dL	0.20	0.38

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

**LAB ADDRESS :**

Govt. UPHC, Kallukulam  
KALLUKULAM, N.K. ROAD  
Thanjavur  
TAMILNADU613001

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*[Signature]*  
**Dr. A.S. MUTHU KUMAR**  
MEDICAL OFFICER  
UPHC, KALLUKULAM,  
THANJAVUR  
9865656519

*[Signature]*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential