



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - OCTOBER 2022



Lab Name Govt. UPHC, Kallukulam  
Constituent Group Chemistry III  
PT item Lyophilized Serum

Lab No 9388  
Date of Result Entered : 29/10/2022  
Date of Report Published : 03/11/2022

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1837	346.52	7.17	24.84	320 mg/dL	-1.07	1.16
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1831	1.36	10.34	0.14	1.1 mg/dL	-1.84	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1783	109.51	6.27	6.87	98 mg/dL	-1.68	0.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

#### LAB ADDRESS :

Govt. UPHC, Kallukulam  
KALLUKULAM, N.K. ROAD  
Thanjavur  
TAMILNADU 613001

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

*Pamela Christudoss*  
Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

*fluv*  
Dr. A.S. MUTHU KUMAR  
MEDICAL OFFICER  
UPHC, KALLUKULAM,  
THANJAVUR.  
9865656519

Homogeneity and Stability of the sample is passed.  
Data in CMC EQAS reports is confidential