



# CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY  
CMC EXTERNAL QUALITY ASSURANCE SCHEME  
MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name **Govt. UGPHC, Rajathany**Lab No **7225**Constituent Group **Chemistry III**Date of Result Entered : **17/11/2022**PT item **Lyophilized Serum**Date of Report Published : **06/12/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik ( Semi automatic / Fully Automatic )	1868	82.95	8.62	7.15	<b>90.7</b> mg/dL	1.08	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik ( Semi automatic / Fully Automatic )	1843	6.98	7.13	0.50	<b>6.97</b> mg/dL	-0.02	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik ( Semi automatic / Fully Automatic )	1767	112.60	5.74	6.46	<b>122.3</b> mg/dL	1.50	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within $\pm 1.01$ to $\pm 2.00$	Good.
Within $\pm 2.01$ to $\pm 2.99$	Accept with caution. Warning Signal.
Beyond $\pm 3.0$	Unacceptable performance. Action Signal.

## LAB ADDRESS :

Govt. UGPHC, Rajathany

THE BLOCK MEDICAL OFFICER, THENI DISTRICT

Rajathany

TAMILNADU625512

Coordinator Contact Details:  
Email: [clinqc@cmcvellore.ac.in](mailto:clinqc@cmcvellore.ac.in)  
Contact Number: 0416-2283102

Dr. Pamela Christudoss  
CMC EQAS Coordinator  
Christian Medical College, Vellore

**Homogeneity and Stability of the sample is passed.**

**Data in CMC EQAS reports is confidential**

**CMC EQAS does not sub contract any components**

**\*\*\*\*\* End of Report \*\*\*\*\***