

HORIBA India Private Limited

246, Okhla Industrial Estate Phase-III,
New Delhi 110020, India
Tel : +91 (11) 4646 5000 / 4669 5001
Fax : +91 (11) 4646 5020
http : //www.horiba.com
CIN : U73100DL2006PTC153232

HIN/MED/2023/236783

10th January 2023

CALIBRATION CERTIFICATE

This is to certify that the hematology analyzer Yumizen H550 bearing serial no: 109YAXH03499 Installed at Lupin Diagnostic Ltd, Burdwan(CAMRI) was calibrated on 10th January 2023.

Calibrator : ABX MINOCAL
Lot No. : CX478
Expiry Date : 05th February 2023

The reports of Blank Cycle, Repeatability and Calibration Values were all found in acceptable range.

Next calibration cycle is due on 10th January 2024



A handwritten signature in blue ink, appearing to be 'Shrish Dixit'.

Shrish Dixit

(Marketing Manager)

For Horiba India Pvt Ltd