

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination
Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score	
SE1	CRP	Negative	Not reported	Negative		Not Applicable			2	2	
SE2	CRP	Positive	24	Positive	31.0011	19.1529	6 to 192	1.6328	-0.4	2	2
SE3	CRP	Negative	Not reported	Positive	20.3592	12.0561	6 to 192	1.0525	Not applicable	2	0

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination
Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score
SE1	RF	Negative	Not reported	Negative		Not Applicable			2	2
SE2	RF	Negative	Not reported	Negative		Not Applicable			2	2
SE3	RF	Negative	Not reported	Negative		Not Applicable			2	2

* As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.