



Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-03000945		Date of Cal. 06.01.2023
Mode: OV-WB		Calibrator Lot No. PLUS0123			Exp.Date:02.05.2023	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	8.92	4.61	13.6	92.4	255	8.92
1.	8.70	4.71	14.6	92.4	252	8.54
2.	8.81	4.74	14.3	92.2	247	8.41
3.	8.72	4.76	14.4	92.1	252	8.52
4.	8.64	4.76	14.4	91.8	245	8.44
5.	8.75	4.77	14.4	91.9	269	8.45
6.	8.65	4.74	14.3	91.6	251	8.63
Mean	8.71	4.75	14.4	92.0	253	8.50
CV(%)	0.731	0.459	0.69	0.30	3.4	0.945
New Cal. Factor (%)	102.39	97.14	94.39	100.45	100.95	
Old Cal. Factors (%)	102.84	93.69	96.16	100.86	97.66	
Old TC1 (%)						103.27
New TC1 (%)						102.53

For Rapid Diagnostic Pvt. Ltd.

N. Ramakrishna Reddy

N. Ramakrishna Reddy
Service Engineer

N. Mahender Reddy

N. Mahender Reddy
Zonal Service Manager




Calibration Certificate

Customer Name	Telangana Diagnostics VIKARABAD
Address	Area Hospital MAHABUBNAGAR -509001
Contact Person	SWATHI
Contact No.	9581111863
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-03000945
Instrument Status	Under Warranty
Calibration Done On	06.01.2023
Next Calibration Due On	05.01.2024

For Rapid Diagnostic Pvt. Ltd.



N. Ramakrishna Reddy
Service Engineer



N. Mahender Reddy
Zonal Service Manager