



CC-3045

AVIAN TEST LAB PVT. LTD.

CIN : U73100AS2018PTC018749



avian

Calibration Certificate

| | | |
|--|-------------------------|-----------------------|
| ATL/F/17 | Page No. | : 1 of 1 |
| Certificate No. : ATL2301-013/02 | Issue Date | : 30-01-2023 |
| ULR No. : CC304523000000235F | Discipline | : MECHANICAL - VOLUME |
| Customer Name & Address: M/s GMCH Medical College and Hospital Guwahati-781032, Assam | SRF No. | : ATL2301013 |
| | SRF Date | : 11-01-2023 |
| | Calibration Date | : 28-01-2023 |
| | Calibration Due Date | : 27-01-2023 |
| | Location Of Calibration | : At Lab |

Details of Instrument Under Calibration:

| | | | |
|---------------------------------|------------------|-----------------------|----------------------------------|
| Description | : Micropipette | ID No. | : GMCH/PPTCT/MP-01 |
| Make | : Handypette | Sr. No. | : 221050639 |
| Range | : 100 to 1000 µl | | |
| Resolution | : 10µl | | |
| Environmental Conditions | | Visual Inspection | : Ok |
| Temperature | : (23±3)°C | Calibration Procedure | : MV-03, As per IS 1997, ISO 385 |
| Humidity | : (50±10)% RH | | |

Detail of Standard Instruments Used :

| Instrument Used | Make& Model | Serial/ID. No. | Valid up to | Certificate No. |
|-----------------------------|-----------------------------|------------------------------------|-------------|-----------------|
| Electronic Weighing Balance | LCCGRADWAG, As 82/220.R2 | 598157, ATLPL/CAL/WS/AS82/220R2 | 03-12-2023 | NC-614 |

Calibration Results:

| Sl. No. | DUC (µl) | Measured Value (µl) | Deviation (µl) | Expanded Uncertainty (Ue) ± (µl) |
|---------|----------|---------------------|----------------|----------------------------------|
| 1 | 100 | 100.00123 | -0.00123 | 1.5 |
| 2 | 200 | 200.00126 | -0.00126 | |
| 3 | 400 | 400.00127 | -0.00127 | |
| 4 | 800 | 800.00129 | -0.00129 | |
| 5 | 1000 | 1000.00131 | -0.00131 | |

Remarks :

- The instrument was received in good condition and was calibrated at lab.
- This certificate pertains only to the item calibrated.
- The calibration results reported in this certificate are valid at the time of and at the stated environmental conditions.
- The calibration interval is determined based on customer's requirements.
- The calibration is traceable to National standards as per traceability details given in the certificate.
- This calibration certificate shall not be reproduced in full, except with prior written approval of CEO, Avian Test Lab Private Limited.,
- This calibration certificate is meant for scientific and industrial purpose only.
- The NABL Symbol is used as per NABL guidelines in NABL-133.
- The Measurement Uncertainty is reported approximately at 95% confidence level with coverage factor $k = 2$.
- The volumetric readings given above are expressed at a reference temperature of 27°C.

DUC*=Device Under Calibration

Calibrated by
(Calibration Engineer)Mr. Vishwas
(Technical Manager)
AUTHORISED SIGNATORY1ST FLOOR, MUKUNDA BHAWAN, SIX MILE, G. S. ROAD, GUWAHATI- 781022, ASSAM
PHONE : +91-9864274915/8486027966, E-MAIL : aviantestlab@gmail.com, Website : www.aviantestlab.com

Scanned with OKEN Scanner