

Date: 15-01-2023
Effective Date: 15-01-2023

Certificate of Calibration

Customer Name: Medicentre Diagnostics Bhilwara

Model : Fully Automated Biochemistry Analyzer EM 360

Serial No. : 60567

Calibration Done Date: 15-01-2023

Next Calibration Due Date : On or Before 15-01-2024

Lab In-charge: Mr. Murli Chipa

This is to certify that the above-mentioned product has been verified of calibration for all biochemistry parameters.

Calibration at site performed by
Nilesh Sharma
Area Service Manager
Ajmer

Technical Service Department
Transasia Bio-Medicals Ltd
Mumbai

Encl: 1. Certificate of Inspection

Date: 15-01-2023
Effective Date: 15-01-2023

Certificate of Inspection

1. Model: Fully Automated Biochemistry Analyzer EM360
2. Serial No.: 60567
3. Calibration Date: 15-01-2023
4. Material used: ERBA NORM , PATH, MULTICAL

By comparing your data to the results of the standard counters in Erba Control, the calibration for all parameters using the measurement standard material was completed. Please refer to the attached sheets for the details.

Technical Service Department
Transasia Bio-Medicals Ltd.
Mumbai

Calibration Data:

1. Instrument Model & Sr. no. : EM360, Sr. no 60567
2. Range of the instrument: Optical density Linear form 0-3.5
3. Date of Calibration - 15, JAN, 2023
4. Next Calibration due on - 15,JAN, 2024

Test Points	Actual	Pass/ Fail
Logic 5 V	4.99V	PASS
Line voltage 220 \pm 10%	225V	PASS
Lamp Voltage 12v	11.9V	PASS
RGT Temperature below 12°C	12°C	PASS
Rotor Temperature 37°C \pm 0.2	37°C	PASS
Sample/ Reagent tray Initialize	OK	PASS
Arm Assembly Initialize	OK	PASS
Rotor Tray Initialize	OK	PASS
Com port	Com 5	PASS
Arm Trough valve	OK	PASS
Arm Probe Clean Valve	Ok	PASS
Waste Pump	OK	PASS
Syringe Initialize	OK	PASS
Dead volume Calibration	OK	PASS
Sample Tube 5-7 ml	350ul	PASS
Sample Cup 2 ml	100ul	PASS
Reagent 50ml Bottle	2000ul	PASS
Reagent 20 ml Bottle	1500ul	PASS
Reagent 5 ml Tube	500ul	PASS

Technical Service Department
Transasia Bio-Medicals Ltd.

Mumbai

7. CALIBRATION DATA

Sr.No.	Filters	Gains	PASS/ FAIL
1	340	79	PASS
2	376	84	PASS
3	415	98	PASS
4	450	77	PASS
5	480	77	PASS
6	505	87	PASS
7	546	88	PASS
8	570	89	PASS
9	600	80	PASS
10	660	94	PASS
11	700	96	PASS
12	750	88	PASS

Technical Service Department
Transasia Bio-Medicals Ltd.

Mumbai