Roche Diagnostics India Pvt. Ltd.

Field Service Report



Case number: CAS-0016432102 Order Number: WO-02112314 | Visit Date: 31/01/2023

Instrument Details Customer details

Instrument/Module: COBAS INTEGRA 400 plus Customer Number: 0052612616

Serial Number: 420523 Customer Name: Mananthavady Diagnostic Centre

Internal Instrument Name: Street Address: Club Kunnu Road,

: 670645 - Manathavady

Additional Details : Ms Jiya

: PM Visit

Purpose of Visit

PM COBAS INTEGRA 400 plus Major

Performed Activities

Carried put routine pm visit, cleaned the instrument, done instrument checks, done instrument calibration, run qc and samples. Instrument working fine

Time Report

Category	Start	End	Hours	Invoice Type
05913616001-Service labour time	31/01/2023 12:25 pm	31/01/2023 4:00 pm	3.57	Warranty

Travel

Value	Hours
Travel	4.42

Customer Parts Used

Part Number	Description	Quantity
05670713001-KIT MAINTENANCE I400/400PLUS	KIT MAINTENANCE I400/400PLUS	1

Field Service Report



Case number: CAS-0016432102 Order Number: WO-02112314 | Visit Date: 31/01/2023

Signature

Customer Signature

Ms Jiya



Roche support Prajil CK



Customer Signature

Roche Signature

The customer acknowledges the service intervention as performed in accordance with Roche recommendations:

- This Service report has been signed by the authorized representative of your organization.
- Discrepancies if any must be intimated within a period of seven (7) days of the receipt of the email, in the absence of which the report will be final and conclusive.
- This report shall be governed by the laws of India and the courts at Mumbai shall have the exclusive jurisdiction to deal with the disputes arising hereunder.

Service Report Version Number: 1