



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY

CMC EXTERNAL QUALITY ASSURANCE SCHEME

MONTHLY SUMMARY REPORT - DECEMBER 2022



PC-1024

Lab Name Govt. Block PHC, Vellakoil
 Constituent Group Chemistry III
 PT item Lyophilized Serum

Lab No 4598
 Date of Result Entered : 19/12/2022
 Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	242.78 mg/dL	0.24	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	3.51 mg/dL	-1.24	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	113.18 mg/dL	-0.28	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Vellakoil
 DHARAPURAM MAIN ORAD, TIRUPUR DISTRICT
 Vellakoil
 TAMILNADU638111

Coordinator Contact Details:
 Email: clinqc@cmcvellore.ac.in
 Contact Number: 0416-2283102

Pamela Christudoss
 Dr. Pamela Christudoss
 CMC EQAS Coordinator
 Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
 Data in CMC EQAS reports is confidential
 CMC EQAS does not sub contract any components
 ***** End of Report *****

14/1/23
 உதவி மருத்துவ அயலவர்
 அரசு குடும்ப சுகாதார நிலையம்,
 வெள்ளையாங்காணி, திருப்பூர் டி.ஓ.,

Dr. J. Jeyapalan
 Dist. Microbiology
 O/o DDHS
 Tpr.