



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - OCTOBER 2022



PC-1024

Lab Name Govt. Block PHC, Vellakoil Lab No 4598
Constituent Group Chemistry III Date of Result Entered : 20/10/2022
PT item Lyophilized Serum Date of Report Published : 03/11/2022

Sl. No.	Analyte	Method Principle Name	Analyzer Name	No. of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1837	346.52	7.17	24.84	317.79 mg/dL	-1.16	1.16
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1831	1.36	10.34	0.14	1.34 mg/dL	-0.14	0.01
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1783	109.51	6.27	6.87	98.69 mg/dL	-1.58	0.33

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. Block PHC, Vellakoil
DHARAPURAM MAIN ORAD, TIRUPUR DISTRICT
Vellakoil
TAMILNADU638111

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

cy
உதவி மருத்துவ அலுவலர்
அரசு ஆரம்ப சுகாதார நிலையம்,
வெள்ளக்கோவில், திருப்பூர் Dt.,