



TBS INDIA

TELEMATIC AND BIOMEDICAL SERVICES PVT. LTD.

CERTIFICATE OF PERFORMANCE FOR CENTRIFUGE

ISSUED TO : ASSAM MEDICAL COLLEGE & HOSPITAL, DIBRUGARH
 CERTIFICATE NO : TBS/CAL/AMCH-DBR/23/0001
 LOCATION/DEPT. : PPTCT DATE : 01.02.2023
 ASSET NO : NA MODEL NO : CLE-129 DM
 MANUFACTURER : COSLAB SERIAL NO : 2020-21/1025

CALIBRATION STANDARD USED

S.No	Asset No	Test Equipment	Manufacturer	Model No	Serial No	Cal. Date	Cal. Due Date
1	TBS/TE/0290	DIGITAL TACHOMETER	EQUINOX	EQ801B	S202194	07.05.22	07.05.23
2	TBS/TE/0187	ELECTRICAL SAFETY ANALYZER	RIGEL	288+	25J-0316	11.05.22	10.05.23

TEST REPORT

S.No	Parameters	Unit	Set Value	Measured Value	Acceptable Range	Remarks
1	SPEED	RPM	1000	1005	± 5%	PASS
			1500	1510	± 5%	PASS
			3000	3005	± 5%	PASS
			3500	3510	± 5%	PASS

Electrical Safety Test Passed as per IEC Standard 60601.1

Calibration Passed: YES NO

Calibration Done Date: **01.02.2023**

Calibration Due Date: **01.02.2024**

Calibrated By: JAGDEEP SINGH

Approved By: _____



The above mentioned instrument has been certified with accuracy guidelines traceable to the National / International Standards

5th Floor, Arden Fair, Opp Benniganahalli Ring Road Flyover, Pai Layout, Near Tin Factory, Old Madras Road,
 Bangalore – 560016, Tel: +91 80 40545050 Fax: +91 80 40545055
 Email: info@tbs-india.com Website: www.tbs-india.com