

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Report No:	PM-CG-2022-06-0483	Date: 18-08-2022	Page 01 of 02
------------	--------------------	------------------	---------------

Customer Details			
Health Facility	Simga CHC (ICTC)		
State	Chhattisgarh	District	Baloda Bazar

Medical Equipment Details			
ME Number	CGZ1BAZCH0110021	Manufacturer	NA
ME Name	Centrifuges, Tabletop(Centrifuge)	Model	NA
Department	Ground Floor - Pathology	Serial No	NA
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	28/06/2022	Due for Calibration	27/06/2023

Environmental Conditions			
Temperature in °C	39°C	Humidity in % Rh	44%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0106	Tachometer	AMPROBE	TACH-10	18070144	08-05-2023

Performance Test						
No	Parameters	Unit	Set (rpm)	Measured (rpm)	Acceptable Range	Remarks
1	Speed	rpm	500	495	475 to 525	Pass
			1000	1025	950 to 1050	Pass
			1500	1555	1425 to 1575	Pass
			2000	2099	1900 to 2100	Pass
			2500	2459	2375 to 2625	Pass

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval.

Calibration Status


Passed

Failed

Calibrated By

Prakash chandan

Authorised Signatory



Arif Jusair