

CALIBRATION REPORT

Refrigerators, Laboratory (Lab refrigerator)



Report No: CAL-CG-2022-08-1304

Date: 20-09-2022

Page 01 of 02

Customer Details			
Health Facility	Nagari CHC		
State	Chhattisgarh	District	Dhamtari

Medical Equipment Details			
ME Number	CGZ1DHACH0460010	Manufacturer	LG
ME Name	Refrigerators, Laboratory(Lab refrigerator)	Model	Gl-b171rphm/2017
Department	1St Floor - Laboratory	Serial No	Na
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	22/08/2022	Due for Calibration	21/08/2023

Environmental Conditions			
Temperature in °C	29°C	Humidity in % Rh	76%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0066	ELECTRICAL SAFETY ANALYZER	RIGEL	288 PLUS	28K-0391	13/12/2022
2	CHC/TE/0088	THERMOMETER	METRAVI	DMT-900	171112502	02-12-2022

Performance Results				
No	Parameter	Measured Value	Acceptable Range	Remarks
1	Temperature	3.9	3.5 to 4.5°C	Pass


CALIBRATION REPORT

Refrigerators, Laboratory (Lab refrigerator)



Conditions of Issue of report

1. The Standards used are traceable to National standards.
2. This report should not be reproduced except in full without written approval from Trimed Solutions India Private Limited.

Calibration Status	Passed <input checked="" type="checkbox"/>	Failed <input type="checkbox"/>
Calibrated By	Authorized Signatory	
Dhal Singh Sahu	 Arif Jaisair*	

