

CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Report No:	CAL-CG-2022-08-1305	Date: 20-09-2022	Page 01 of 02
------------	---------------------	------------------	---------------

Customer Details			
Health Facility	Nagari CHC		
State	Chhattisgarh	District	Dhamtari

Medical Equipment Details			
ME Number	CGZ1DHACH0460011	Manufacturer	Remi
ME Name	Centrifuges, Tabletop(Centrifuge)	Model	R-8C
Department	1St Floor - Pathology	Serial No	DLH-5777
Calibrated at	On Site	Condition of ME	Good
Date of Calibration	22/08/2022	Due for Calibration	21/08/2023

Environmental Conditions			
Temperature in °C	29°C	Humidity in % Rh	76%

Details of Master Equipment Used for Calibration						
No	Equipment ID	Description	Manufacturer	Model	Serial No.	Calibration Validity
1	CHC/TE/0106	Tachometer	AMPROBE	TACH-10	18070144	08-05-2023

Performance Test						
No	Parameters	Unit	Set (rpm)	Measured (rpm)	Acceptable Range	Remarks
1	Speed	rpm	500	510	475 to 525	Pass
			1000	1000	950 to 1050	Pass
			1500	1500	1425 to 1575	Pass
			2500	2600	2375 to 2625	Pass
			3500	3600	3375 to 3625	Pass



CALIBRATION REPORT

Centrifuges, Tabletop (Centrifuge)



Conditions of issue of report

- 1 The Standards used are traceable to National standards
- 2 This report should not be reproduced except in full without written approval

Calibration Status

Passed

Failed

Calibrated By

Ghat Singh Saha

Authorized Signatory

