

CENTRIFUGE CALIBRATION REPORT

DEPARTMENT	INSTRUMENT NAME	INSTRUMENT ID	MODEL NO	REPORT NO
Accession	Centrifuge	ACC/001	RBC	CAL/CFG/2 Jan-23/02
Cal Done Date	18012023	23	Cal Due Date	18072023
Ownership	SRL <input checked="" type="checkbox"/>	LOAN <input type="checkbox"/>	RR <input type="checkbox"/>	
Acceptable rpm/time limit	± 5 %	Operating Range	From 100 rpm To 4000 rpm	

Sr. No.	Set Value	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.
1	1000	1006.4	2.51	5.935	6.325
2	2000	2007.0	2.55	6.635	6.338
3	4000	4013.4	2.302	13.935	6.262

Sr. No.	Set Value	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.
1	60	60.32	0.13	0.086	0.164
2	300	300.48	0.164	0.089	0.247
3	600	600.88	0.13	0.293	0.372

Sr. No.	Set Value °c	Measured Value (Avg. of 5 reading)	SD	Error	Exp. Uncert.

The reported expanded uncertainty is based on a standard uncertainty multiplied by a coverage factor k=2, providing the level of confidence of approx. 95%.

Calibration status	Passed <input checked="" type="checkbox"/>	Failed <input type="checkbox"/>
--------------------	--	---------------------------------

TEST EQUIPMENT NAME	Test equipment ID	Traceable Report No
TACHOMETER	ENG/DTM-01	MTL/SRL2/R02/01-23
TIMER	ENG/SW-01	MTL/SRL2/R06/01-23
THERMOMETER		

The above mentioned instrument has been calibrated using standards with accuracy guidelines traceable to the National Institute of Standards & Technology (NIST).

Engineer's Name	<i>Mirajul Islam</i>	Report Reviewed by	
Signature/Date	<i>18/1/23</i>	Signature/Date	<i>D. Saha</i>

SECTION HEAD SIGN:

Doc No.: ENG QC02(316)	Centrifuge Calibration Report			Page 1 of 1
Issue No.: 01	Issue Date: 01.12.2021	Amend No.: 00	Amend Date: 00	
Issued By: Mr. Surendra Kumar Gupta (DGM-QA)				<i>Sup 5</i>



MEASURE TECHNO LAB
 2, B.T. ROAD (JAYANTI CINEMA COMPLEX), BARRACKPORE,
 KOLKATA - 700120, W.B.



Phone: 033 - 2215 - 0032, 2215 - 9687, 8100875519, Mobile: 9831190974,
 LAB:- 8100143376, E-mail: measuretechno@yahoo.co.in

CALIBRATION CERTIFICATE OF DIGITAL TACHOMETER

CALIBRATION CERTIFICATE NO.: MTL / SRL / R02 / 01 - 23

ULR - CC254523000001407F

Page: 1 of 1

1.0 Service Request No.: MTL / 14 / 01 / 22 - 23

**TRACEABILITY DOCUMENT FOR
 REFERENCE USE FOR
 CALIBRATION.**

1.1 Issued to: M/s. SRL Limited,
 P.S. SriJan Tech Park, Ground Floor,
 DN - 52, Sector V, Salt Lake,
 Kolkata - 700091, W.B.

1.2 Description & Identification of Item to be Calibrated:

a) Name:	Digital Tachometer	b) Code No.:	ENG / DTM - 01
c) Sl. No.:	AB.16577	d) Make:	Lutron
e) Model / Type:	DT - 2236	f) Range:	5 rpm to 99999 rpm
g) Sensor:	N.A.	h) Resolution:	0.1 rpm upto 999.9 rpm then 1 rpm
i) End User:	Engineering Department	j) Accuracy:	N.S.
k) Calibration done at:	On Site / <input checked="" type="checkbox"/> In House		

1.3 Date of receipt of Item : 13-01-23

1.4 Physical Condition of DUC : OK

1.5 Date of calibration : 14-01-23

1.6 Recommended date of next calibration : 14-01-24

1.7 Date of Issue : 16-01-23

1.8 Environmental Conditions During Calibration: Temperature: 20 °C ± 2 °C
 Humidity: 50 % RH ± 10 % RH

1.9 Method of Calibration: SOP / RPM / 01

2.0 Traceability :
 a) Standards used for calibration are traceable to National standards through NABL Accredited Laboratory.
 b) The following standards / Equipment have been used.
 i) Digital Tachometer Cal. Certificate No. JRPM - CCTR - A&S - 2022 - 0025 (JRPM, Chennai) (Cal. Date: 09/07/22, Due Date: 08/07/23)

2.1 Result :
 Mechanical Calibration

Sl. No.	Parameter/ Range	Nominal Value on DUC rpm	Std. Value Average of Five Readings rpm	Error rpm	Measurement Expanded Uncertainty ± rpm
1.		100.1	100.11	-0.01	0.28 upto 100 rpm then 1.28 upto 10000 rpm then 14 upto 30000 rpm then 10.43
2.		499.9	499.85	0.05	
3.		1000	1000.5	-0.5	
4.	Rotation	4999	4999.6	-0.6	
5.	5 rpm to 99999 rpm	9987	9986	1	
6.		19999	19995	4	
7.		29988	29992	-4	
8.		59983	59980	3	
9.		89485	89487	-2	

Remarks: i) This result has an expanded uncertainty with a coverage factor k=2 at approximately 95% confidence level.
 ii) The calibration certificate issued for this instrument is to be used for scientific or industrial purposes only.
 iii) Error = DUC Reading - Standard Reading.
 DUC - Device Under Calibration N.S. - Not Specified N.A. - Not Applicable
 Std. - Standard rpm - Rotation Per Minute

Opinions and Interpretations	
Calibrated	✓
Limited Use	Accepted / Valid for use
	Rejected / Out of use

Calibration Engineer
 Measure Techno Lab
 K.K. Singh
 Calibration Engineer

Checked / Approved