



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2022



Lab Name Govt. UGPHC, S Goodaloor

Lab No 10622

Constituent Group Chemistry III

Date of Result Entered : 20/11/2022

PT item Lyophilized Serum

Date of Report Published : 06/12/2022

Sl No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1868	82.95	8.62	7.15	93 mg/dL	1.41	0.33
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1843	6.98	7.13	0.50	7.3 mg/dL	0.64	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1767	112.60	5.74	6.46	115 mg/dL	0.37	0.31

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

Govt. UGPHC, S Goodaloor
CLINICAL LABORATORY , TIRUVANNAMALAI District .,
S Goodaloor
TAMILNADU606753

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2293102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

M. S. S. S.
Medical Officer
Govt. Primary Health Cent. -
Se.Gudalur-606 753
Thandarampattu Taluk.