



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - DECEMBER 2022



Lab Name Govt. UGPHC, S Goodaloor Lab No 10622
Constituent Group Chemistry III Date of Result Entered : 19/12/2022
PT Item Lyophilized Serum Date of Report Published : 05/01/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE III	GOD-POD III	Robonik (Semi automatic / Fully Automatic)	1844	239.20	6.25	14.96	246 mg/dL	0.45	0.70
2	CREATININE III	Jaffes Kinetic - Alkaline Picrate III	Robonik (Semi automatic / Fully Automatic)	1839	3.91	8.24	0.32	3.6 mg/dL	-0.96	0.02
3	CHOLESTROL III	CHOD-PAP III	Robonik (Semi automatic / Fully Automatic)	1766	114.94	5.41	6.21	119 mg/dL	0.65	0.30

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :
Govt. UGPHC, S Goodaloor
CLINICAL LABORATORY , TIRUVANNAMALAI District .,
S Goodaloor
TAMILNADU606753

Coordinator Contact Details:
Email:clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components
***** End of Report *****

M. S. S.
Medical Officer
Govt. Primary Health Centre
Se.Gudalur-606 753
Thandarampattu Taluk.