

The Scientific Engineering Works

No.4, Second Cross, Chinthamani Nagar, Sababu Colony, Coimbatore - 641 038.

Phone: 9422 34402505 | Mobile: 94433 00205

E-mail: tsewcebe@gmail.com | apthangadurai@gmail.com

www.thescientificengineeringworks.com

GSTIN : 33AELPT1783J1Z5

CALIBRATION CERTIFICATE

Customer

M/S UGPHC

Certificate No : TSEW-550

Date of calibration : 07.11.2022

Se., Gudalur

Calibration due Date : 06.11.2023

DETAILS OF THE TEST INSTRUMENT

Description	SEMIAUTOANALYZER
Make	ROBONIK
Model	PRIESTEST TOUCH
Serial Number	ATCD-1050519RBK
Range	-
Deviation	-
Accuracy	-
Condition of the Equipment receipt	Functional

CALIBRATION ENVIRONMENTS

Temperature	26°C ± 2°C
Relative Humidity	35 % to 65 %

DETAILS OF MASTER INSTRUMENT USED FOR CALIBRATION

Description	Make	Certified by	Certificate No
HBAIC REAGENT	-	-	-

CALIBRATION DATA

S.NO	Test Instrument Reading	Master Instrument Reading	Deviation	Remark
1	4.0%	4.0%	-	OK
2	5.0%	5.0%	-	OK
3	6.0%	6.0%	-	OK

Remarks :

1. The above calibration readings were taken as per customer requirements.
2. Overall Deviation is within Specified Accuracy. Hence Instrument is OK.
3. Maximum Tolerance Permissible With in limit.

CALIBRATED BY:


SERVICE ENGINEER
(A.MUTHUKUMAR)

