

The Scientific Engineering Works

No 4, Second Cross, Chinthamani Nagar, Seibaba Colony, Coimbatore - 641 038

Phone: 9422 24482585 | Mobile: 94433 88285

E-mail: tsewco@gmail.com | apthangadurai@gmail.com

www.thescientificengineeringworks.com

GSTIN : 33AELPT1763J125

CALIBRATION CERTIFICATE

Customer

M/S UGPHC

Se, Gudalur

Certificate No : TSEW -551

Date of calibration : 07.11.2022

Calibration due Date : 06.11.2023

DETAILS OF THE TEST INSTRUMENT

Description	CELLCOUNTER
Make	SYSMEX
Model	85804
Serial Number	-
Range	-
Deviation	-
Accuracy	-
Condition of the Equipment receipt	Functional

CALIBRATION ENVIRONMENTS

Temperature	26°C +2° C
Relative Humidity	35 % to 65 %

DETAILS OF MASTER INSTRUMENT USED FOR CALIBRATION

Description	Make	Certified by	Certificate No
-	-	-	-

CALIBRATION DATA

S.NO	Test Instrument Reading	Master Instrument Reading	Deviation	Remark
1	RBC OK	-	-	OK
2	WBC OK	-	-	OK
3	PLATELET OK	-	-	OK

Remarks :

1. The above calibration readings were taken as per customer requirements.
2. Overall Deviation is within Specified Accuracy. Hence Instrument is OK.
3. Maximum Tolerance Permissible With in limit.

CALIBRATED BY:


SERVICE ENGINEER
(A.MUTHUKUMAR)

