

ARUNACHALA SCIENTIFIC CO

16A, APS Building, 2nd Cross, Om Ganesh Nagar, Vadavalli, Coimbatore - 641 041.
E-mail : arunachalascientifico@gmail.com www.arunachalascientifico.com

☎: 90477 36633
94442 46633

GSTIN : 33BTGPP6365K1Z8

CALIBRATION CERTIFICATE

CUSTOMER NAME & ADDRESS

UPGRADED GOVT.PRIMARY HEALTH CENTER,
SE- GUDALORE.

CERTIFICATE NO :ASC/C/2022/580
CALIBRATION DATE : 03.11.2022
NEXT CALIBRATION DATE:02.11.2023

Description : LABOMED MICROSCOPE
Model No : VISION 2000
S.I.NO : 130761837
LOCATION : LABORATORY
EYE PIECES : 10X.W.F.ONE PAIR
OBJECTIVE : 4X, 10X, 40X HIGH POWER AND 100X OIL
IMMERSION
LIGHT SOURCE : HALOGEN LAMP
ILLUMINATION CONTROL : LIGHT SOURCE LOW & HIGH
POWER IS WORKING IN GOOD CONDITION
MECHANICAL STAGE : FRONT AND BACK & LEFT AND
RIGHT MOVING IS GOOD
COARSE : GOOD, STAGE UP AND DOWN
MOVEMENT IS IN GOOD CONDITION
FINE TUNER : 0.001mm GOOD CONDITION

FINAL TEST AND OBSERVATION OF THE MICROSCOPE IS WORKING IN
GOOD CONDITION

AUTHORISED SIGNATORY



P.PALANI
Quality/Technical Manager

