



CUSTOMER SERVICE & CALIBRATION
REPORT

MUTHU SURGICALS

GST NO : 33DDEPM1487D1ZQ

UDYAM REGISTRATION-TN- 16-0005072

101,ANNAI COMPLEX,NEW BUSTANT, PERAMBALUR,621212.

CSR NO.

DATE: 15.08.2022

Customer Name:

THE MEDICAL OFFICER

Address:

MURUKKANKUDI PHC

CENTRIFUGE

State: tamil I nadu

ZIP Code: 621133

Status of Call: **Warranty or AMC**
(Please circle)

Call Logged on: 08-12-2022

NATURE OF PROBLEM

Problem Reported:

- 1.REGULATER PROBLEM
- 2.COIL PROBLEM
- 3.WIRE PROBLEM
- 4.FIXED ANGLE ROTOR
- 5.SWING OUT ROTOR

System Status: Operable

Equipment Type: CENTERIFUGE

Make: INDIA

Model LABOTECH

Serial No. BDI-152

Call Reported by:

Date:10-12-2022

Time: 1.PM

NEXT DUE DATE: 10-12-2023

SERVICE DETAILS

Service Rendered:

- 1.COIL CHECK 220V AC 60HZ
- 2.REGULATER PRESSUR AND GAGE CHECK 3500 RPM
- 3.WIRE POWER INPUD AND OUTPUT CHECK



<p>Engineer s Remarks:</p> <p>THE LABORATORY CENTRIFUGE YEARLY SERVICE MAINTANCE COMPULSORY</p>	<p>Status after Service: (Please circle) Complete Working solution provided</p> <p>RESULTS:</p> <p>THE LABORATORY CENTRIFUGE CHECK THE ALL COMPOUNDS CONTISION IS GOOD CALIBIRATION SUCCESSF</p>
-------------------------------------------------------------------------------------------------	----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------


Spares Replaced:

Events: (Date & Time)	Start of Service: 10.12.2022	End of service: 10-12-2022
------------------------	---------------------------------	----------------------------

CUSTOMER FEEDBACK

Remarks:

Name	Designation:	Phone/Fax: 8056356534
------	--------------	-----------------------

<p>Engineer Signature  For MUTHU SURGICALS Name: NS.MUTHAZHAGANBE., (Bio medical eng) Date :10-12-2022</p>	<p>Customer Signature & Seal:</p>
-------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------	---------------------------------------