

testing: 11/01/2023

Specimen Log in Register for Proficiency Testing and EQA Panel Samples Received or Sent by SRL/NRE

RETESTING
 Name of the ICTC/PPPT/Blood Bank/PPP-ICTC: Cottage Hospital Chicalim
 Quarterly Period from: JAN 2023 - March 2023

SITE ID	SAMPLE ID	DATE RECEIVED	TYPE & QUALITY OF SAMPLE *	DETAILS OF HIV TEST KIT USED AND RESULTS OBTAINED BY THE SENDING SITE			DETAILS OF HIV TEST KIT USED AND RESULTS OBTAINED BY SRL/NRE			FINAL RESULT AT SRL/NRE	DATE RESULT SENT TO SITE	COMMENTS
				HIV TEST 1	HIV TEST 2	HIV TEST 3	HIV TEST 1	HIV TEST 2	HIV TEST 3			
E08				Name: Combid Lot No: 4000027211 Expiry Date: 2024/06/11	Name: Meliscree Lot No: MI102131 Expiry Date: 2023/09	Name: Trideo Lot No: MI02235 Expiry Date: 2023/01	Name: Combid R Lot No: 4000027211 Expiry Date: 11/06/2024	Name: Meliscree Lot No: MI102131 Expiry Date: 09/2023	Name: Trideo HIV Lot No: MI022135 Expiry Date: 01/2023			
A	PWSAICTC GA S600001023 0021	02/11/23	Good quality	Non-Reactive (NR)	-	-	NR	-	-	Negative		100% Matching
B	PWSAICTC GA S600001023 0042	04/11/23	Sufficient volume	Non-Reactive (NR)	-	-	NR	-	-	Negative		100% Matching
C	GCSAICTC GA S600001023 0084	06/11/23		Non-Reactive (NR)	-	-	NR	-	-	Negative		100% Matching
A	PWSAICTC GA S600001023 0024	02/11/23		Reactive (R)	Reactive HIV-1 (R)	Reactive HIV-1 (R)	R (CI=5.0)	R (HIV-1)	R (HIV-1)	Positive for HIV-1		

* Type & quality of sample = Positive, Negative, Indeterminate, Good Quality, Sufficient Volume
 No. of Discordant results: 0
 Reasons for Discordant results: -
 Action Taken: -

R = Reactive
 NR = Non Reactive

Dr. Gurleen Kaur
 Associate Professor
 Department of Microbiology
 Goa Medical College
 Bambolim-Goa

HEALTH OFFICER
 SUB DISTRICT HOSPITAL
 CHICALIM, VASCO - GOA