

KIRLOSKAR TECHNOLOGIES (P) LTD.

(Quality Assessment Services)

B-58, 1st Floor, Defence Colony, Bhisham Pitamah Marg, New Delhi-110024, INDIA
Phone : 011-24331368, 24331369, 41403967; E-mail : gianender.verma@kirloskarmedical.com



Technology from the
frontiers of medical science

Calibration Certificate

Page 1 of 2

Certificate Number:	KTPL/OD/CAL/2022-02/02605		ULR No.	N/A	
Calibration Date:	18-02-2022	Next Recommended Cal. Date:	18-02-2023	Date of Issue:	19-02-2022
DUC Receipt Date:	18-02-2022	SRF No.	2021-22/0567	SRF Date:	18-02-2022
Issued to (Customer Details):-	M/S.DHH Nayagada Main Road, Odisha - 752069			Chalan/ Reference :-	---
Equipment Location:				Chalan/ Reference Date :-	---
Details of Equipment /Device Under Calibration(DUC):-					
Name of equipment:	Centrifuge M/C 8Tube	Equipment ID:	N/A		
Make:	Remi	Range/ Size:	Upto 3500 RPM		
Model/ Type:		Least Count:	---		
Serial No.		Accuracy / Tolerance:	Not provide by Customer		
KTPL Bar Code.	162197	Operational Manual:	Not provide by Customer		
Condition of DUC:	Satisfactory	Calibrated at (Lab/Site):	Site		
Environmental Conditions During Calibration:					
Temperature in (°C):	25 ± 4 °C		Relative Humidity in (%):-	60 ± 20 %	
Calibration Reference standard :					
KIPI/CP48-01-00		IEC 60601-1/62353, SANAS TR 45-01 & 02			
Detail of Reference standard Instruments used for Calibration:					
Description	Make/Model.	Serial No.	Certificate No.	Cal. Validity	
Digital Tachometer	Line Seiki/TM 5010	18Z 0040V	AACPL/14543F	5-Sep-2022	
Visual Inspection of Device Under Calibration :- OK					
Calibration/Performance result :-					
Sr. No.	Parameter	Unit of Measurement	DUC knob set	STD Value	Expended Uncertainty of measurement (± %)
1)	Rotation	RPM	1	1148.7	1.6
			2	1556.4	1.6
			3	2160.0	1.6
			4	2526.2	1.6
			5	3054.4	1.6

A Singh
Calibrated By: (Ashish Kr Singh)

Satyapaal Singh
Satyapaal Singh
Sr. Branch Service Manager
Kirloskar Technologies Pvt. Ltd.



Ashish Kr. Singh
Approved and Released By: (Ashish Kr. Singh).