

Calibration Certificate

ULR No.:	CC30962300000836F		
Certificate No.:	CCTPL/DF/0081/02	Field:	Thermal
SRF No.:	50/23	Date of Receipt:	08/02/2023
SRF Date:	08/02/2023	Date of Calibration:	08/02/2023
Company Name & Address	M/s. Genesis Path Lab	Recommended Due Date.:	07/02/2024
	240, Sector-05, Vasundhara, Ghaziabad-201012	(As per agreed by customer)	
		Certificate Issue Date:	09/02/2023

DUC* Details

Instrument Name:	Digital Thermometer With Sensor	Instrument Sr. No.:	-----
Make:	Anmol	Instrument Id No.:	-----
Model No.:	ST-9269	Location:	Genesis Path Lab
Range:	-50 to 300 °C	DUC Condition:	OK
Readability:	0.1 °C	Calibration Performed at:	At Site

Standard Equipments Used(Traceable to National Standard)

Sr. No	Instrument Name	Cert. No	Cal. By	Calibration Due Date
1	SPRT With Super DAQ	TL/022/1095.1.1	Tempsens	30/09/2023

Environment Condition	Temperature	Humidity	Reference Standard	Calibration Procedure No.
		26.2°C	54 %RH	DKD-R-5-7


Calibration Result

Sr. No.	Nominal Temperature (°C)	Standard Reading (°C)	DUC* Reading (°C)	Deviation (°C)	Expanded Uncertainty ±(°C)
1	2	2.015	2	-0.015	1.6
2	4	4.022	4	-0.022	
3	6	6.034	6	-0.034	
4	8	8.046	8	-0.046	

Note : DUC* : Device Under Calibration

Uncertainty of measurement at approx 95% Confidence Level with coverage factor k = 2.

Calibrated By : 
Nishant Kumar
Calibration Engineer



Approved By : 
Sandeep Sharma
Director

Note. 1.This Certificate only to the particular instrument submitted for calibration.
2.This Calibration Result reported in this certificate valid at the time of and under the stated condition of measurement.
3.This particular certificate can not be reproduce except in full, without prior permission of CCTPL.

--- End of Certificate ---

Page 1 of 1



0120-4105146 | 9810703533 | 9899899577



info@calyss.in

