

SPARK

SCIENTIFIC PRIVATE LTD

CIN No.: U51909TG2017FTC115492

TO,
Redcliffe Lifetech Private Limited,
F82, Ground Floor, Patel Nagar,
Gwalior, MP - 474011

CALIBRATION PROCEDURE

Equipment Name: Easy Electrolytes
Equipment Serial No. 0122041006

We Spark Scientific Pvt Ltd hereby confirm that Calibration procedure of Easy Electrolytes instrument that this device is firmly calibrated by manufacturer and does not required any regular calibration.

M.V. Mahesh Kumar



Name: M.V. Mahesh Kumar
Desg: Director

Spark Scientific Private Limited

Plot No. 5A/10-11, IDA Nacharam, Road No.1, Uppal Mandal, Hyderabad - 500 076. T.S. India.
Tel.: +91 40 27172293 / 5088 | Email: info@sparksci.in | Website: www.sparksci.in



Performance Qualifications for EASY ELECTROLYTES Analyzer (Serial No: 0122041006)

Customer Name: REDCLIFFE LIFETECH PVT LTD
F-82, Patel Nagar City Centre in front of Osho Medication Centre, Gwalior

Calibration Parameters

Verified that all the control parameters were within the acceptable limits

and in range. Verified for precision and found that control values were within acceptable CV.

System Certification:

Study data has determined that the system described in this document either meets the necessary criteria outlined in this performance qualification protocol.


The system is ready for specific usage.

Protocol performed by:

Ahmad
Name: *Ishtiyag Ahmad*

Designation: *Asst Service Manager*

Performance Qualification



done by: Designation:

M^{rs}. Mamoj
Lab supervisor

QC RESULTS
LAST CONTROL
QC LEVEL 1
JAN-08
Na⁺ 140.9
K⁺ 3.70
FEB-09-23; 18:35

QC RESULTS
LAST CONTROL
QC LEVEL 1
JAN-08
Na⁺ 140.9
K⁺ 3.70
Cl⁻ 103.6
FEB-09-23; 18:35

Handwritten signature and date: 09/02/23

FEB-09-23; 18:46

QC LEVEL 3
LOT# 00021321

Na⁺ 117.7 mmol
K⁺ 2.97 mmol
Cl⁻ 95.1

Handwritten signature and date: 09/02/23

