



PROFICIENCY TESTING REPORT
ISHTM-AIIMS EXTERNAL QUALITY ASSURANCE PROGRAMME
 NABL accredited program as per ISO/IEC 17043:2010 standard
 Organized By Department of Hematology, AIIMS, New Delhi-110029



Duration of stability testing - minimum upto 8 days at ambient temp. after dispatch of specimens

EQAP CODE No. : 4355

Distribution No.: 156-L

Month/Year: July/2022

Instrument ID: YUMIZEN H550

Name & Contact No. of PT Co-ordinator: Dr. Seema Tyagi (Prof.), Hematology, AIIMS, Delhi,
 Tel: 9013085730 , E-Mail : accuracy2000@gmail.com

Date of issue & status of the report: 27-09-2022[Final].

CBC and Retic Assessment

Test Parameters	S.No.	Among Lab (Accuracy Testing)						Within Lab (Precision Testing)			
		Your Result 1	Your Result 2	Your Results Sum of 2 Value	Consensus result sum of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score	Yours Results Diff. of 2 Values	Consensus Result Diff. of 2 values (Assigned Value)	Uncertainty of Assigned Values	Z Score
WBC x10 ³ /µl	1	6.1	4.75	10.85	12.18	0.0990	-0.45	1.35	0.15	0.0120	7.36
RBC x10 ⁶ /µl	1	4.4	4.36	8.76	8.69	0.0140	0.18	0.04	0.05	0.0040	-0.15
Hb g/dl	1	12.5	12.5	25	24.8	0.0290	0.25	0	0.1	0.0100	-0.45
HCT%	1	39.5	39.3	78.8	79.6	0.2150	-0.14	0.2	0.4	0.0290	-0.34
MCV-fl	1	90.1	89.7	179.8	183.3	0.3570	-0.35	0.4	0.4	0.0270	0.00
MCH-Pg	1	28.7	28.5	57.2	57.3	0.0800	-0.05	0.2	0.3	0.0200	-0.34
MCHC-g/dl	1	31.9	31.7	63.6	62.4	0.1610	0.28	0.2	0.3	0.0230	-0.27
Plt. x10 ³ /µl	1	190	187	377	348	1.70	0.65	3	7	0.47	-0.45
Retic %	2										

P.S . Assesment

YOUR REPORT		CONSENSUS REPORT
DLC%	3	Nrbcs=5 , Poly=44 L=1, E=3, Mono/Promono=0 , B1=0 P.M.=1, Mye=33, Meta=18, Other=
RBC Morphology	3	Predominantly: Normocytic/Normochromic; Moderate: Anisocytosis, hypochromia, Microcytosis; Mild: Macrocytosis, Poikilocytosis
Diagnosis	3	Chronic Myeloid Leukemia (Chronic Phase)

COMBINED DATA VALUES OF TOTAL PARTICIPANTS

Test parameters	S.No.	Total participants covered in the current dist. 156--L	Total No. responded	% of Labs with Z Score 0-2		% of Labs with Z Score 2-3		% of Labs with Z Score >3	
				Among labs	Within lab	Among labs	Within lab	Among labs	Within lab
WBC x10³/µl	1	318	306	92.16	82.35	5.56	5.88	2.28	11.77
RBC x10⁶/µl	1	318	318	87.11	83.65	5.35	4.4	7.54	11.95
Hb g/dl	1	318	318	87.74	84.91	3.14	5.35	9.12	9.74
HCT%	1	318	307	92.18	89.25	4.56	5.21	3.26	5.54
MCV-fl	1	318	307	91.21	91.86	5.54	3.58	3.25	4.56
MCH-Pg	1	318	307	89.25	85.67	4.89	3.58	5.86	10.75
MCHC-g/dl	1	318	307	90.88	86.97	5.86	5.21	3.26	7.82
Plt. x10³/µl	1	318	307	91.86	92.18	5.86	2.61	2.28	5.21
ReticCount%	2	318	218	91.28	83.94	3.67	13.3	5.05	2.76
PS Assessment	3	318	216	Satisfactory :93.4%, Borderline Sat. :2.83%, Unsatisfactory :3.77%					

***Comments:**

1). **Among Lab (EQA) : Results acceptable.**

2). **Within Lab (IQA) : Difference in the CBC measurement values for WBC unacceptable, may be due to random/human error.**

Note-1: EQA (External Quality Assurance) : Your Performance among various of participating labs in PT, to determine the accuracy of your results.

IQA (Internal Quality Assurance) : Your Performance of comparison of two consecutive measurement values within your lab to test the precision of your autoanalyzer.

Note-2: Z score among & within lab were calculated, as per to ISO/IEC 13528:2015 standard. Z score among lab (EQA)= (Your Result Sum of two values - Consensus Result sum of two values)/(Normalised IQR)

Z score within lab (IQA)= (Your Result Difference of two values - Consensus Result difference of two values)/(Normalised IQR)

IQR = Quartile 3 - Quartile 1 of participant data, Normalised IQR = 0.7413 x IQR

Note-3: Z score 0 to ±2: Acceptable, Z score ±2 to ±3 :Warning Signal, Z score > ±3 : Unacceptable [As per ISO/IEC 13528:2015 standard]

Note-4: Z score value between "0 to ±2" are texted in green colour. Z score value between "±2 to ±3" are texted in orange colour. Z score value > ±3 are texted in red colour.

Note-5: Homogeneity and stability testing of PT sample were done as per ISO 13528:2015 standard. To pass homogeneity test, between sample SD (Ss) should be smaller than the check value (0.3*SDPA). To pass the stability test, average difference in measurement values of first and last day sample ($\bar{x}-\bar{y}$) should be smaller than the check value (0.3*SDPA).

Note-6: ISHTM-AIIMS-EQAP does not subcontract any task of its scheme

Note-7: Participants are free to use methods/analyzer of their own choice.

Note-8: Proficiency testing (PT) samples are sent quarterly to each participant.

Note-9: All the necessary details regarding design and implementation of PT, are provided in the instruction sheet as well as on programme's website www.ishtmaiimseqap.com.

Note 10: Reports are kept confidential.

Report authorized by,



Dr. Seema Tyagi (Prof.)

PT Co-ordinator: ISHTM-AIIMS-EQAP

Department of Hematology, AIIMS, New Delhi

-----End Of Report-----