

31.01.2023

CALIBRATION CERTIFICATE

EPSILYTE Electrolyte Analyser Na/K/Cl

Serial No: 1601134 installed at Pullickal Medical Trust Hospital, Kuzuppilly

Tested for the following with Calibrator pack Lot no: 20220415; Exp: 14.10.2023

Slope of Na (Sodium)	- 49.7 (Instrument Range 27-67)
Slope of K (Potassium)	- 45.0 (Instrument Range 27-67)
Slope of Cl (Chloride)	- 50.1 (Instrument Range 27-60)

Ran Quality Control and found results satisfactory.

Na	-	145.6	Range (Mean 145 \pm 2.0)
K	-	5.08	Range (Mean 5.0 \pm 0.1)
Cl	-	106.3	Range (Mean 105 \pm 2)

Next Calibration due on 30th July 2023

For Epsilon Diagnostics



Unni Karthik

Senior Sale/ Service Engineer