



CHRISTIAN MEDICAL COLLEGE

DEPARTMENT OF CLINICAL BIOCHEMISTRY
CMC EXTERNAL QUALITY ASSURANCE SCHEME
MONTHLY SUMMARY REPORT - NOVEMBER 2022



PC-1024

Lab Name **JAVIX LIFE PVT LTD** Lab No **16513**
Constituent Group **Chemistry II** Date of Result Entered : **19/11/2022**
PT item **Lyophilized Serum** Date of Report Published : **06/12/2022**

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE II	GOD-POD II	Transasia / Erba	840	186.15	10.35	19.27	183.5 mg/dL	-0.14	1.33
2	UREA II	Urease UV / GLDH II	Transasia / Erba	700	51.95	11.66	6.06	51.6 mg/dL	-0.06	0.46
3	CREATININE II	Jaffes End point II	Any Analyser (Automation / Semi Automation)	667	2.53	12.67	0.32	2.82 mg/dL	0.90	0.02
4	T.BILIRUBIN II	Diazonium Salt (Colorimetric) / Jendrassik II	Transasia / Erba	756	2.92	13.06	0.38	1.94 mg/dL	-2.57	0.03
5	T-PROTEIN II	Biuret - Colorimetric II	Transasia / Erba	724	5.75	8.92	0.51	4.78 g/dL	-1.89	0.04
6	ALBUMIN II	BCG - Colorimetric II	Transasia / Erba	713	3.30	8.67	0.29	3.26 g/dL	-0.14	0.02
7	URIC ACID II	Enzymatic / Uricase Colorimetric II	Transasia / Erba	772	5.66	14.64	0.83	5.2 mg/dL	-0.55	0.06
8	CHOLESTROL II	CHOD-PAP II	Transasia / Erba	822	110.52	12.60	13.93	107.5 mg/dL	-0.22	0.97
9	TRIGLYCERIDE II	GPO-PAP / Enzymatic Colorimetric / End Point II	Transasia / Erba	760	160.62	13.16	21.14	155 mg/dL	-0.27	1.53

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ± 1.01 to ± 2.00	Good.
Within ± 2.01 to ± 2.99	Accept with caution. Warning Signal.
Beyond ± 3.0	Unacceptable performance. Action Signal.

LAB ADDRESS :

JAVIX LIFE PVT LTD
B2-702, PACIFIC GOLF ESTATE, SAHASTRADHARA ROAD
DEHRADUN
UTTARAKHAND248001

Coordinator Contact Details:
Email: clinqc@cmcvellore.ac.in
Contact Number: 0416-2283102

Pamela Christudoss
Dr. Pamela Christudoss
CMC EQAS Coordinator
Christian Medical College, Vellore

Homogeneity and Stability of the sample is passed.
Data in CMC EQAS reports is confidential
CMC EQAS does not sub contract any components

***** End of Report *****