



PC-1033

111th IAMM EQAS Microbiology: Bacteriology/ Serology

CMC MICRO EQAS

Department of Clinical Microbiology, Christian Medical College, Vellore-632004, Tamil Nadu

Email: egas@cmcvellore.ac.in, Twitter: @microcgas, Phone: +91-416-2282588

JULY 2022

111th EQAS EVALUATION REPORT

MEMBER ID:

M 1 2 1 1

JULY 2022 / SEROLOGY

Test method employed for detection C-reactive protein (CRP) at your lab: Latex Agglutination

Peer group (n) = 229

Please refer the attached evaluation format for details on the criteria for evaluation

Marks Obtained: 10/10 (100%)

Parameter	Your Result	Your Value (mg/L)	Intended Result	Robust Mean	Robust SD	Range (mg/L)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	CRP	Negative	<6	Negative	Not Applicable						2	2
SE2	CRP	Positive	48	Positive	31.0011	19.1529	6 to 192	1.6328	0.9	2	2	
SE3	CRP	Positive	24	Positive	20.3592	12.0561	6 to 192	1.0525	0.3	2	2	

JULY 2022 / SEROLOGY

Test method employed for detection Rheumatoid Factor (RF) at your lab: Latex Agglutination

Peer group (n) = 406

Please refer the attached evaluation format for details on the criteria for evaluation

Parameter	Your Result	Your Value (IU/mL)	Intended Result	Robust Mean	Robust SD	Range (IU/mL)	Uncertainty of Assigned value	Z & Z'	Max Marks	Your Score		
SE1	RF	Negative	<8	Negative	Not Applicable						2	2
SE2	RF	Negative	<8	Negative	Not Applicable						2	2
SE3	RF	Positive	16	Negative	NOT EVALUATED*							

*As >30% participants scored <70% on this exercise, it has been excluded from the final evaluation.

Disclaimer:

This is a confidential document and subject to the rules of confidentiality as described by the ISO 17043:2010 standard.

MEMBER ID:

M 1 2 1 1

SE1	SE2	SE3	Marks obtained	
4	4	2	10	100%
4	4	2	Maximum marks = 10	

Dr. Rani Diana Sahni
Scientific Co-ordinator

Dr. John A Jude Prakash
Quality Manager

Dr. V. Balaji
PT Co-ordinator

Report Dispatch Date: 15.11.2022

***** End of Report *****