



Calibrator Calibration

Instrument: 5 Part Haematology Analyzer		Model: BC -6200		Serial No.: TW-04001028		Date of Cal. 13.07.2022
Mode: OV-WB		Calibrator Lot No. PLUS0722			Exp.Date:05.08.2022	
Parameter	WBC	RBC	HGB	MCV	PLT	WBC-D
TARGET	9.05	4.59	13.6	92.5	255	9.05
1.	8.91	4.92	14.4	93.8	251	8.83
2.	8.93	4.87	14.3	93.8	235	8.90
3.	8.98	4.91	14.3	93.6	237	8.80
4.	8.79	4.93	14.4	93.5	252	8.99
5.	9.10	4.85	14.3	93.1	247	8.72
6.	8.88	4.88	14.3	93.5	249	8.65
Mean	8.93	4.89	14.3	93.5	245	8.81
CV(%)	1.162	0.643	0.29	0.28	3.0	1.391
New Cal. Factor (%)	101.32	93.81	94.95	100.03	103.97	
Old Cal. Factors (%)	103.31	91.51	95.84	98.88	92.84	
Old TC1 (%)						101.11
New TC1 (%)						101.34

For Rapid Diagnostic Pvt. Ltd.


N. Ramakrishna Reddy
 Service Engineer


N. Mahender Reddy
 Zonal Service Manager



Calibration Certificate

Customer Name	Telangana Diagnostic Hub Sircilla
Address	OPP. MRO office, SIRCILLA-505301
Contact Person	Mr. Ramprasad
Contact No.	9959013841
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-04001028
Instrument Status	Under Warranty
Calibration Done On	13.07.2022
Next Calibration Due On	12.07.2023

For Rapid Diagnostic Pvt. Ltd.



N. Ramakrishna Reddy
Service Engineer



N. Mahender Reddy
Zonal service Manager