




## Calibrator Calibration

<b>Instrument:</b> 5 Part Haematology Analyzer		<b>Model:</b> BC -6200		<b>Serial No.:</b> TW-02000921		<b>Date of Cal.</b> 05.08.2022
<b>Mode: OV-WB</b>		<b>Calibrator Lot No. PLUS0722</b>			<b>Exp.Date:05.08.2022</b>	
<b>Parameter</b>	<b>WBC</b>	<b>RBC</b>	<b>HGB</b>	<b>MCV</b>	<b>PLT</b>	<b>WBC-D</b>
<b>TARGET</b>	9.05	4.59	13.6	92.5	255	9.05
1.	8.46	4.71	14.4	94.8	261	8.64
2.	8.47	4.63	14.4	94.7	261	8.33
3.	8.62	4.63	14.4	94.6	252	8.48
4.	8.66	4.70	14.3	94.2	246	8.46
5.	8.52	4.64	14.3	94.6	247	8.45
6.	8.68	4.68	14.3	94.1	249	8.58
<b>Mean</b>	8.87	4.66	14.4	94.5	253	8.49
<b>CV( %)</b>	1.135	0.792	0.20	0.30	2.8	1.278
<b>New Cal. Factor (%)</b>	105.62	98.43	92.98	97.89	100.95	
<b>Old Cal. Factors (%)</b>	103.98	91.38	94.63	101.27	93.10	
<b>Old TC1 (%)</b>						101.16
<b>New TC1 (%)</b>						100.95

**For Rapid Diagnostic Pvt. Ltd.**



**G. Nagaraju**  
Service Engineer



**N. Mahender Reddy**  
Zonal Service Manager



## Calibration Certificate

<b>Customer Name</b>	<b>Telangana Diagnostics NIRMAL</b>
Address	Area Hospital NIRMAL, NIRMAL-504106
Contact Person	Mr. MADHU SUDHAN
Contact No.	7075824647
Instrument	5 Part Haematology Analyzer
Model	BC 6200
Serial No.	TW-02000921
Instrument Status	Under Warranty
Calibration Done On	05.08.2022
Next Calibration Due On	04.08.2023

**For Rapid Diagnostic Pvt. Ltd.**



**G. Nagaraju**  
Service Engineer



**N. Mahender Reddy**  
Zonal Service Manager