

View Lab Monthly Summary

Lab Name TELANGANA DIAGNOSTICS
 Lab No 18012
 Month January
 Year 2023
 Constituent Group Chemistry I

[Details About Robust Analysis](#)
[Detail About Monthly Summary](#)
[Detail about SDI](#)

Click on the analyte to view Graphical Data

All Analyser Result Print

Date of Result Entered : 23/01/2023

Date of Report Published : 09/02/2023

Sl.No	Analyte	Method / Principle Name	Analyzer Name	No of Participants	DV	Participants		Your Value	SDI	U
						CV	SD			
1	GLUCOSE	Hexokinase	Any Analyser (Automation / Semi Automation)	157	174.21	4.68	8.15	178 mg/dL	0.46	1.30
2	UREA	Urease UV / GLDH	Any Analyser (Automation / Semi Automation)	588	31.57	10.65	3.36	24 mg/dL	-2.25	0.28
3	CREATININE	Jaffes Kinetic - Alkaline picrate	Any Analyser (Automation / Semi Automation)	767	1.40	10.99	0.15	1.4 mg/dL	0.00	0.01
4	T.BILIRUBIN	Diazonium salt (Colorimetric) / Jendrassik	Any Analyser (Automation / Semi Automation)	835	3.43	15.36	0.53	3.9 mg/dL	0.89	0.04
5	T-PROTEIN	Biuret - Colorimetric	Any Analyser (Automation / Semi Automation)	956	5.25	9.77	0.51	5.3 g/dL	0.10	0.03
6	ALBUMIN	BCG - colorimetric	Any Analyser (Automation / Semi Automation)	674	3.31	8.11	0.27	3.3 g/dL	-0.04	0.02
7	CALCIUM	Arsenazo III	Any Analyser (Automation / Semi Automation)	865	9.49	6.54	0.62	8.9 mg/dL	-0.95	0.04
8	URIC ACID	Enzymatic / Uricase Colorimetric	Any Analyser (Automation / Semi Automation)	939	4.83	16.02	0.77	4.3 mg/dL	-0.68	0.05
9	CHOLESTEROL	CHOD-PAP	Any Analyser (Automation / Semi Automation)	974	113.44	5.89	6.68	110 mg/dL	-0.51	0.43
10	TRIGLYCERIDE	GPO-PAP / Enzymatic Colorimetric / End Point	Any Analyser (Automation / Semi Automation)	935	109.13	7.82	8.54	123 mg/dL	1.62	0.56
11	HDL	Direct method / Enzymatic colorimetric	Any Analyser (Automation / Semi Automation)	780	30.17	20.45	6.17	30 mg/dL	-0.03	0.44
12	AST	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	967	50.29	10.00	5.03	48 U/L	-0.46	0.32
13	ALT	UV kinetic(with & without PLP (P-5-P))	Any Analyser (Automation / Semi Automation)	958	57.78	13.00	7.51	60 U/L	0.30	0.49
14	ALP	PNP AMP kinetic	Any Analyser (Automation / Semi Automation)	793	101.50	10.81	10.97	97 U/L	-0.41	0.78
15	AMYLASE	CNPG3	Any Analyser (Automation / Semi Automation)	439	43.41	18.75	8.14	52 U/L	1.06	0.78

SDI Range	Interpretation
Within -1.00 to +1.00	Excellent.
Within ±1.01 to ±2.00	Good.
Within ±2.01 to ±2.99	Accept with caution. Warning Signal.
Beyond ±3.0	Unacceptable performance. Action Signal.

Links

[Modify Address](#)
[Yearly Summary](#)

[Modify Methods](#)
[Change Password](#)

[Submit Lab Results](#)
[Home](#)

[View Lab Monthly Summary](#)
[Sign-out](#)